MCH-1000





The Quarterly Maintenance Procedure is a reprocessing cycle which cleans and disinfects the external surfaces and water path of the unit. To ensure proper operation of the MCH, perform the Initial Operation Check found in the MCH Operator/Service Manual at least quarterly. Additionally, ensure all accessories and modules are reprocessed at the same time as the MCH unit.

Date	Device S/N	Disinfectant Used/ Volume	Cleaned By	Notes	Next Cleaning Date